



<b>GROUP:</b> -		<b>SITE:</b> St Johns	<b>PERFORMED:</b> 09/12/2020 21:08:49 mike (Administrator)		
<b>Alternate ID</b>	JDOE	<b>Height</b>	186.0cm; 73.2in	<b>Acceptability</b>	Good Session
<b>Sex at Birth</b>	Male	<b>BMI</b>	27.7	<b>Device Name</b>	Pneumotrac Model-6800
<b>Age</b>	40.90	<b>Smoking</b>	Non Smoker	<b>Device Serial</b>	PN06028
<b>Date of Birth</b>	15/01/1980	<b>Predicted Set</b>	GLI Caucasian	<b>Accuracy</b>	04/12/2020 16:46:00
<b>Population</b>	Caucasian	<b>Correction Factor</b>	100 %	<b>Temperature</b>	17.5°C
<b>Weight</b>	96.0kg; 211.6lb	<b>Num of Tests</b>	5		



Values at BTPS		Base					
Parameter	Best	LLN	Z-Score	%Pred	Test 1 21:08:49	Test 5 21:17:36	Test 3 21:15:49
<b>FVC (L)</b>	<b>6.34</b>	4.64	0.69	108.75%	6.34	6.29	6.02
<b>FEV1 (L)</b>	<b>4.61</b>	3.67	-0.05	99.35%	4.57	4.61	4.53
<b>FEV1/FVC</b>	<b>0.73</b>	0.70	-1.15	-	0.72	0.73	0.75
<b>FEV1/VC</b>	<b>0.74</b>	0.70	-0.99	-	0.73	0.74	0.73
<b>FET (s)</b>	<b>11.44</b>	-	-	-	11.44	10.63	7.94
<b>PEF (L/s)</b>	<b>8.90</b>	8.38	-1.30	81.50%	8.86	8.84	7.79

Session Quality and Repeatability										
FVC Rep:	FEV1 Rep:	FVC Grade	FEV1 Grade	Slow Start	Cough in 1st sec	FEV1 Rejected	No Flow Plateau	FIVC-FVC	EOFE Unacceptable	FVC Rejected
0.05L (0.79%)	0.04L (0.87%)	A	A	0 blow(s)	0 blow(s)	0 blow(s)	1 blow(s)	0 blow(s)	1 blow(s)	0 blow(s)

Base VC							
Parameter	Best	LLN	Pred	%Pred	Test 2 21:06:55	Test 3 21:07:38	Test 1 21:05:36
<b>VC (L)</b>	<b>6.24</b>	4.64	5.83	107.03	6.24	6.20	6.12
<b>IC (L)</b>	<b>3.43</b>	3.01	4.16	82.53	3.64	3.35	3.31

Interpretations	
<b>User</b>	-
<b>GOLD ATS/ERS</b>	Computer interpretations cannot be relied upon for diagnosis. Normal Spirometry.
<b>Estimated Lung Age</b>	-
<b>Session Notes</b>	-